

Tender Book annex / ملحق دفتر الشروط - AHF-Procurement -2023-13			C.2-PR234 ICU Equipment	
Commercial Name / Ticari Ad / الاسم التجاري		Return Quotation to /Teklifin Teslimi / مستلم العرض		
no (رقم)	اسم المواد ووصفها ITEMS DESCRIPTION/SPECIFICATION Malzemenin Adı ve Özelliği	المواصفات الفنية Specification Teknik Özellikler	الملاحظات Notes Notlar	العلامة التجارية & بلد المنشأ Brand Details & Country of Origin
1	ICU bed	ICU bed with bed matterss and air matteress- Turkish		
2	Air Conditioner 1 TON	Inverter air condition / Energy saving-Turkish		
3	Air Conditioner 2 TON	Inverter air condition / Energy saving-Turkish		
4	Defibrillation and Cardioversion and shock divice	<p>Should be a Low Energy Biphasic defibrillator monitor with Recorder, having capability to deliver shocks from 2 Joules to 200 Joules. Should monitor ECG through paddles, pads and monitoring electrodes and Defibrillate through pads and paddles. Should compensate for body impedance for a range of 25 to 150 ohms Should have a built in 50 mm strip printer Should have charging time of less than 5 seconds for maximum energy.Should have High resolution more than 8 inch Colour display for viewing monitoring parameters like ECG, SpO2, NIBP and etCO2 with 4 waveform capability of 4 seconds. Should have event summary facility for recording and printing at least 55 events. Should have a battery capable of usage for at least 5 hours of monitoring. capable of printing Reports on Event summary, configuration, self test, battery capacity etc. Should have facility for self test/check before usage and set up function.Should have facility to monitor parameters like SpO2, NIBP and etCO2 along with non invasive pacing (Demand & Fixed mode) facility. Should be able to upgrade the defibrillator for 12 lead ECG monitoring and ECG transmission. German origin</p>		

5	Suction device	Portable , equipped with two bottles, with strong pull, powered by electricity 220-Italian -Turkish		
---	----------------	---	--	--